

## Supplementary Materials

Supplementary Table S1. Instrument parameters for Inductively Coupled Plasma Mass Spectrometer.

Parameter	Value	Unit
Nebulizer Gas Flow	0.69	L/min
ICP RF Power	1100	W
Lenz Voltage	6	V
Analog Stage Voltage	-2300	V
Pulse Stage Voltage	1600	V