

```
R(reflections)= 0.0230( 1380)      wR2(reflections)=
S = 1.048                          0.0612( 1473)
Npar= 143
```

The following ALERTS were generated. Each ALERT has the format

test-name_ALERT_alert-type_alert-level.

Click on the hyperlinks for more details of the test.



Alert level C

PLAT029_ALERT_3_C	_diffn_measured_fraction_theta_full value Low .	0.971	Why?
PLAT094_ALERT_2_C	Ratio of Maximum / Minimum Residual Density	2.54	Report
PLAT911_ALERT_3_C	Missing FCF Refl Between Thmin & STh/L= 0.595	28	Report
PLAT971_ALERT_2_C	Check Calcd Resid. Dens. 0.92Ang From La1	1.68	eA-3



Alert level G

ABSMU01_ALERT_1_G	Calculation of _exptl_absorpt_correction_mu		
	not performed for this radiation type.		
PLAT003_ALERT_2_G	Number of Uiso or Uij Restrained non-H Atoms ...	4	Report
PLAT004_ALERT_5_G	Polymeric Structure Found with Maximum Dimension	2	Info
PLAT007_ALERT_5_G	Number of Unrefined Donor-H Atoms	2	Report
PLAT128_ALERT_4_G	Alternate Setting for Input Space Group C2/c	12/a	Note
PLAT178_ALERT_4_G	The CIF-Embedded .res File Contains SIMU Records	1	Report
PLAT300_ALERT_4_G	Atom Site Occupancy of O6 Constrained at	0.5	Check
PLAT300_ALERT_4_G	Atom Site Occupancy of C8 Constrained at	0.5	Check
PLAT300_ALERT_4_G	Atom Site Occupancy of C9 Constrained at	0.5	Check
PLAT300_ALERT_4_G	Atom Site Occupancy of C10 Constrained at	0.5	Check
PLAT300_ALERT_4_G	Atom Site Occupancy of H8 Constrained at	0.5	Check
PLAT300_ALERT_4_G	Atom Site Occupancy of H9 Constrained at	0.5	Check
PLAT300_ALERT_4_G	Atom Site Occupancy of H10 Constrained at	0.5	Check
PLAT301_ALERT_3_G	Main Residue Disorder(Resd 1)	14%	Note
PLAT398_ALERT_2_G	Deviating C-O-C Angle From 120 for O3 .	106.2	Degree
PLAT767_ALERT_4_G	INS Embedded LIST 6 Instruction Should be LIST 4		Please Check
PLAT789_ALERT_4_G	Atoms with Negative _atom_site_disorder_group #	7	Check
PLAT794_ALERT_5_G	Tentative Bond Valency for La1 (III) .	3.28	Info
PLAT860_ALERT_3_G	Number of Least-Squares Restraints	18	Note
PLAT883_ALERT_1_G	No Info/Value for _atom_sites_solution_primary .		Please Do !
PLAT909_ALERT_3_G	Percentage of I>2sig(I) Data at Theta(Max) Still	85%	Note
PLAT978_ALERT_2_G	Number C-C Bonds with Positive Residual Density.	3	Info

0 **ALERT level A** = Most likely a serious problem - resolve or explain
0 **ALERT level B** = A potentially serious problem, consider carefully
4 **ALERT level C** = Check. Ensure it is not caused by an omission or oversight
22 **ALERT level G** = General information/check it is not something unexpected

2 ALERT type 1 CIF construction/syntax error, inconsistent or missing data
5 ALERT type 2 Indicator that the structure model may be wrong or deficient
5 ALERT type 3 Indicator that the structure quality may be low
11 ALERT type 4 Improvement, methodology, query or suggestion
3 ALERT type 5 Informative message, check

It is advisable to attempt to resolve as many as possible of the alerts in all categories. Often the minor alerts point to easily fixed oversights, errors and omissions in your CIF or refinement strategy, so attention to these fine details can be worthwhile. In order to resolve some of the more serious problems it may be necessary to carry out additional measurements or structure refinements. However, the purpose of your study may justify the reported deviations and the more serious of these should normally be commented upon in the discussion or experimental section of a paper or in the "special_details" fields of the CIF. checkCIF was carefully designed to identify outliers and unusual parameters, but every test has its limitations and alerts that are not important in a particular case may appear. Conversely, the absence of alerts does not guarantee there are no aspects of the results needing attention. It is up to the individual to critically assess their own results and, if necessary, seek expert advice.

Publication of your CIF in IUCr journals

A basic structural check has been run on your CIF. These basic checks will be run on all CIFs submitted for publication in IUCr journals (*Acta Crystallographica*, *Journal of Applied Crystallography*, *Journal of Synchrotron Radiation*); however, if you intend to submit to *Acta Crystallographica Section C* or *E* or *IUCrData*, you should make sure that full publication checks are run on the final version of your CIF prior to submission.

Publication of your CIF in other journals

Please refer to the *Notes for Authors* of the relevant journal for any special instructions relating to CIF submission.

