

Supplementary Materials:

Hydrophobic Paper-Based SERS Sensor by Gold Nanoparticles Arranged on Graphene Oxide Flakes

Dong-Jin Lee¹ and Dae Yu Kim^{1,2*}

¹ Inha Research Institute for Aerospace Medicine, Inha University, Incheon 22212, Republic of Korea

² Department of Electrical Engineering, College of Engineering, Inha University, Incheon 22212, Republic of Korea

* Correspondence: dyukim@inha.ac.kr; Tel.: +82-032-860-7394

1. Contact angle measurements




Figure 1. (a) The representative shape of water droplets on h-paper substrate. (b) The contact angle measurements of h-paper substrates with different PDMS fraction.

2. Au NPs counting




Figure 2. Particle counting process. Original image is converted to binary image, and a threshold range is manually set. The Au NPs are automatically counted using ImageJ software.

3. Reproducibility




Figure S1. (a) Photographic image of the proposed AuNPs@GO/h-paper SERS substrates. (b) Raman intensity at 612 cm⁻¹ of 10⁻³ M R6G measured by 15 Au NPs@GO/h-paper SERS substrates. The relative standard deviation is about 11.3%.

4. Reliability




Figure 4. Comparison of the Raman intensity of AuNPs@GO/h-paper SERS substrates that kept for 0 and 15 days.

5. Raman characteristic peak of R6G molecule

Table 1. Raman characteristic peak of R6G molecule.

Peak wavenumber (cm⁻¹)	Assignment
613	In plane xanthene ring deformations; Out of plane xanthene ring deformations
775	Out of plane C-H bend; In plane xanthene ring deformations
1184	In plane xanthene ring deformations; In plane C-H bend; In plane N-H bend
1312	In plane xanthene ring breath; In plane N-H bend; CH ₂ wag
1364	Xanthene ring stretch; In plane C-H bend
1512	Xanthene ring stretch; C-N stretch; C-H bend; N-H bend
1577	Xanthene ring stretch; in plane N-H bend
1651	Xanthene ring stretch; in plane C-H bend