

## Supplementary tables

### Microdissected tissue vs tissue slices - A comparative study of two tumor explant models cultured on-chip and off-chip.

Dina Dorigiv<sup>1,2</sup>, Kayla Simeone<sup>1,3</sup>, Laudine Communal<sup>1</sup>, Jennifer Kendall-Dupont<sup>1</sup>, Amélie St-Georges-Robillard<sup>1,4</sup>, Benjamin Péant<sup>1</sup>, Euridice Carmona<sup>1</sup>, Anne-Marie Mes-Masson<sup>1,3</sup> and Thomas Gervais<sup>1,2,4 \*</sup>

Table S1 dimensions of devices

Parameter	value		
Channel height (mm)	MDT chip	0.9	
	Tissue slice chip	Top layer	0.9
		Bottom layer	1.1
Channel width (mm)	MDT chip	1.1	
	Tissue slice chip	Top layer	1.1
		Bottom layer	1.1
Channel length	MDT chip	6	
	Tissue slice chip	Top layer	23
		Bottom layer	18
Tissue diameter (mm)	MDT	0.5	
	Tissue slice	3	
Tissue thickness (mm)	0.35		
Tissue chamber dimension (mm)	MDT well (cubical)	0.7	
	Tissue slice (cylindrical)	Diameter	7
		Height	2.5
	96 well-plate (Corning ®-cylindrical)	Diameter	6.4
		Height	17

Table S2 tissue uptake parameters and diffusion properties

Parameter	value		
Diffusion constant of glucose (cm <sup>2</sup> /s)	Tissue	2.7x10 <sup>-6</sup>	
	Medium	9.27x10 <sup>-5</sup>	
Diffusion constant of oxygen (cm <sup>2</sup> /s)	Tissue	1.8x10 <sup>-5</sup>	
	Medium	2.6x10 <sup>-5</sup>	
	PDMS	3.4x10 <sup>-5</sup>	
Saturation concentration (mM)	Oxygen	Tissue	1.02
		medium	0.21
		PDMS	1.43
	Glucose	11	
Partition coefficient	PDMS-Medium	0.15	
	Medium-Tissue	4.8	
Maximum cellular uptake rate (mM/S)	Oxygen	2.07	
	Glucose	1.09	
Michaelis-Menten constant (mM)	Oxygen	4.63x10 <sup>-3</sup>	
	Glucose	4x10 <sup>-2</sup>	