



Electrochemical Sensor Based on Ni–Co Layered Double Hydroxide Hollow Nanostructures for Ultrasensitive Detection of Sumatriptan and Naproxen

Hadi Beitollahi ^{1,*}, Zahra Dourandish ², Somayeh Tajik ³, Fatemeh Sharifi ³ and Peyman Mohammadzadeh Jahani ⁴

¹ Environment Department, Institute of Science and High Technology and Environmental Sciences, Graduate University of Advanced Technology, Kerman, 7631885356, Iran

² Department of Chemistry, Faculty of Science, Shahid Bahonar University of Kerman, Kerman 76175-133, Iran

³ Research Center of Tropical and Infectious Diseases, Kerman University of Medical Sciences, Kerman 7616913555, Iran

⁴ School of Medicine, Bam University of Medical Sciences, Bam 7661771967, Iran

* Correspondence: h.beitollahi@yahoo.com

Table S1. Selectivity results for 50.0 μ M sumatriptan in the presence of other interferences.

Species	Current (μ A)
50.0 μ M sumatriptan	5.5
50.0 μ M sumatriptan+500-fold of Na ⁺	5.55
50.0 μ M sumatriptan+500-fold of Mg ²⁺	5.57
50.0 μ M sumatriptan+500-fold of Ca ²⁺	5.47
50.0 μ M sumatriptan+500-fold of NH ₄ ⁺	5.53
50.0 μ M sumatriptan+300-fold of fructose	5.60
50.0 μ M sumatriptan+300-fold of glucose	5.63
50.0 μ M sumatriptan+300-fold of lactose	5.61
50.0 μ M sumatriptan+100-fold of histidine	5.42
50.0 μ M sumatriptan+100-fold of phenyl alanine	5.44
50.0 μ M sumatriptan+100-fold of methionine	5.43
50.0 μ M sumatriptan+100-fold of cysteine	5.40
50.0 μ M sumatriptan+20-fold of levodopa	5.60
50.0 μ M sumatriptan+20-fold of uric acid	5.65