



Supplementary Material 3. A clean arthroscopy frame was taken and augmented with a red overlay, where the transparency varies dynamically between 10% and 90%. The artificial images were then used as test inputs for the proposed algorithm.