

Abstract

Optical Fiber Exhaled Breath Sensor Based on Lossy Mode Resonance Using a Graphene Oxide Sensitive Coating [†]

Pedro Sanchez ^{1,2,*}, Miguel Hernaez ², Carlos Ruiz Zamarreño ^{1,3}, Francisco Javier Arregui ^{1,3}, Andrew G. Mayes ² and Sonia Melendi-Espina ²

¹ Department IEE, Universidad Pública de Navarra, 31006 Navarra, Pamplona, Spain; carlos.ruiz@unavarra.es (C.R.Z.); parregui@unavarra.es (F.J.A.)

² Norwich Research Park, University of East Anglia, Norwich NR4 7TJ, UK; M.Hernaez@uea.ac.uk (M.H.); andrew.mayes@uea.ac.uk (A.G.M.); S.Melendi-Espina@uea.ac.uk (S.M.-E.)

³ Institute of Smart Cities, 31006 Navarra, Pamplona, Spain

* Correspondence: pedro.sanchez@unavarra.es

[†] Presented at the 5th International Symposium on Sensor Science (I3S 2017), Barcelona, Spain, 27–29 September 2017.

Published: 6 December 2017

Optical fiber sensors (OFS) have attracted increasing attention due to their benefits over traditional sensors, such as small size, biocompatibility, remote sensing ability or safety in flammable environments. Among the different existing configurations of OFS, those based on electromagnetic resonances are very popular as they are reliable, robust and very sensitive. In particular, sensors based on lossy mode resonance (LMR) are very interesting as a wide range of materials, including metal oxides and polymers, can support them and they do not require specific equipment to tune the optical polarization.

Graphene-based materials such as graphene oxide (GO) or reduced graphene oxide (rGO) have become the most explored materials since Novoselov and Geim achieved its isolation in 2004. Their superior properties, such as high surface area or extreme sensitivity to the external environment, make them ideal candidates for the fabrication of the sensitive coatings required by LMR-based sensors.

In this work, the fabrication and characterization of a small and portable exhaled breath LMR-based OFS using GO as sensitive coating is presented. Refractive index changes have been detected showing a fast repetitive behaviour with a response time of 150 ms from inhalation to exhalation and a high average sensitivity of 410 nm/RIU.



© 2017 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>).